



NEUTRON MEASUREMENTS PRODUCT WITH A ULTRASONIC POLYVALENT PIEZONUCLEARE REACTOR

Steriwave Startec Ltd



The Company

Our core competencies are in design and realization of industrial prototypes of devices with ultrasonic sonotrode.


Board of Directors:

Franco Selleri, Francesco Fucilla, Bruno Denantes,

Adrian J. Eyre, Mike Goddard, Mike Goodrich: Company Secretary

Roberto Mignani - Chief Scientist

Fabio Cardone - Senior Scientific Advisor



Piezonuclear Reactor

This polyvalent reactor is capable of producing piezionuclear reactions on both liquid and solid materials.


Operating conditions of the reactor:

- Subcritical, low production of neutrons;
- Critical, production of pulses of neutrons equal to that published in the scientific literature;
- Supercritical, producing neutrons at levels higher than those published in scientific literature.

The **measurements** were performed with:


- spectrometer Microspec 2 produced by BTI in energy spectra of neutrons produced in subcritical conditions. First measure in historic order of spectra of neutrons produced by reactions piezonuclear.
- neutron detector PDS 100 G-GN/ID produced by Mirion-MGP. Measurement of neutron pulses as a function of elapsed time during the reactor operation in critical condition.

S. Ambrogio R-1-S Project



Measurement of the counting rate of neutron and gamma dose rate.


Background measurement, duration of measurement 5 minutes.



Measurement of the counting rate of neutron and gamma dose rate.

Visit our application to a bar of iron of mass 500 g of 20 kHz ultrasound 19 Watt, 5 minutes duration of the measure. Are evident daily emissions (Burst) of neutrons corresponding to areas of damage the image of the central area of the bar.

S. Ambrogio R-1-S Project




14 March 2009

Neutron spectrum measurements in the laboratory located at Startec in Brugherio (MI).

Measurements taken from Radtech at our laboratories.

S. Ambrogio R-1-S Project




14 March 2009

Neutron spectrum measurements at 20 kHz with ultrasonic power absorbed from the external network 19w duration of the measure 60 minutes.

Material reaction used: Iron

Measurements taken from Radtech at our laboratories.

S. Ambrogio R-1-S Project



14 March 2009

Neutron spectrum measurements at 20 kHz with ultrasonic power absorbed from the external network 19w duration of the measure 60 minutes.

Material reaction used: Iron


Measurements taken from Radtech at our laboratories.




S. Ambrogio R-1-S Project

14 Marzo 2009


Neutron spectrum measurements.



Misura dello Spettro di Neutroni del fondo del laboratorio




Materiale di reazione utilizzato Ferro in polvere sinterizzato e compresso.



Materiale di reazione utilizzato Ferro.

Measurements taken from Radtech at our laboratories.

● ● ● Iron bar subjected to ultrasonic



Iron bar subjected to ultrasound of 20 kHz with power of 19 watts:

- a) iron bar before the application of ultrasound
- b) iron bar after 5 minutes of ultrasound. 6 zones of damage are visible by emission of pulses of neutron flux estimated $10 / 8 - 10 / 10$ neutrons/cm² sec
- c) magnification of one of the damaged regions from the issuance of the neutron pulses

Physical and Chemical Analysis Calibration

Project 1

11/24/2009 4:02:43 PM



Spectrum processing :
No peaks omitted

Processing option : All elements analyzed (Normalised)
Number of iterations = 4

Standard :

C CaCO3 1-Jun-1999 12:00 AM
Cl KCl 1-Jun-1999 12:00 AM
Fe Fe 1-Jun-1999 12:00 AM
Co Co 1-Jun-1999 12:00 AM

| Element | Weight% | Atomic% |
|---------|---------|---------|
| C K | 32.15 | 69.30 |
| Cl K | 1.74 | 1.27 |
| Fe K | 16.21 | 7.51 |
| Co K | 49.90 | 21.92 |
| Totals | 100.00 | |



Physical and Chemical Analysis Normal Sample

Project 1

11/24/2009 4:39:00 PM



Spectrum processing :
No peaks omitted

Processing option : All elements analyzed (Normalised)
Number of iterations = 4

Standard :

C CaCO3 1-Jun-1999 12:00 AM
Si SiO2 1-Jun-1999 12:00 AM
Mn Mn 1-Jun-1999 12:00 AM
Fe Fe 1-Jun-1999 12:00 AM
Dy DyF3 1-Jun-1999 12:00 AM
W W 1-Jun-1999 12:00 AM

| Element | Weight% | Atomic% |
|---------|---------|---------|
| C K | 2.43 | 10.69 |
| Si K | 0.31 | 0.57 |
| Mn K | 0.70 | 0.67 |
| Fe K | 91.41 | 86.41 |
| Dy L | 4.69 | 1.52 |
| W M | 0.47 | 0.13 |
| Totals | 100.00 | |



INCA

Comment: campione normale 1 K(alpha) ferro = 6.4KeV
Tempo acquisizione 120 s
Ingrandimento 3000x

Physical and Chemical Analysis Normal Sample

Project 1

11/24/2009 4:44:48 PM



Spectrum processing :
No peaks omitted

Processing option : All elements analyzed (Normalised)
Number of iterations = 3

Standard :

C CaCO3 1-Jun-1999 12:00 AM
Si SiO2 1-Jun-1999 12:00 AM
Cr Cr 1-Jun-1999 12:00 AM
Mn Mn 1-Jun-1999 12:00 AM
Fe Fe 1-Jun-1999 12:00 AM
Dy DyF3 1-Jun-1999 12:00 AM

| Element | Weight% | Atomic% |
|---------|---------|---------|
| C K | 2.24 | 9.85 |
| Si K | 0.36 | 0.68 |
| Cr K | 0.19 | 0.19 |
| Mn K | 0.68 | 0.66 |
| Fe K | 92.10 | 87.19 |
| Dy L | 4.44 | 1.44 |
| Totals | 100.00 | |



Comment: campione normale 2 K(alpha) ferro = 6.4KeV
Tempo acquisizione 120 s
Ingrandimento 3000x

INCA

Steriwave Startec Ltd

Physical and Chemical Analysis Normal Sample

Project 1

11/24/2009 4:49:09 PM



Spectrum processing :
No peaks omitted

Processing option : All elements analyzed (Normalised)
Number of iterations = 3

Standard :

C CaCO3 1-Jun-1999 12:00 AM
Si SiO2 1-Jun-1999 12:00 AM
Cr Cr 1-Jun-1999 12:00 AM
Mn Mn 1-Jun-1999 12:00 AM
Fe Fe 1-Jun-1999 12:00 AM
Dy DyF3 1-Jun-1999 12:00 AM
W W 1-Jun-1999 12:00 AM

| Element | Weight% | Atomic% |
|---------|---------|---------|
| C K | 2.37 | 10.43 |
| Si K | 0.21 | 0.39 |
| Cr K | 0.18 | 0.19 |
| Mn K | 0.66 | 0.63 |
| Fe K | 91.92 | 86.87 |
| Dy L | 4.12 | 1.34 |
| W M | 0.53 | 0.15 |
| Totals | 100.00 | |



Comment: campione normale 3 K(alpha) ferro = 6.4KeV
Tempo acquisizione 120 s
Ingrandimento 3000x

Inca

Steriwave Startec Ltd

Physical and Chemical Analysis Ultrasonic Sample

Project 1

11/24/2009 4:18:14 PM



Spectrum processing :
No peaks omitted

Processing option : All elements analyzed (Normalised)
Number of iterations = 6

Standard :

C CaCO3 1-Jun-1999 12:00 AM
O SiO2 1-Jun-1999 12:00 AM
Na Albite 1-Jun-1999 12:00 AM
Si SiO2 1-Jun-1999 12:00 AM
S FeS2 1-Jun-1999 12:00 AM
Cl KCl 1-Jun-1999 12:00 AM
K MAD-10 Feldspar 1-Jun-1999 12:00 AM
Mn Mn 1-Jun-1999 12:00 AM
Fe Fe 1-Jun-1999 12:00 AM

| Element | Weight% | Atomic% |
|---------|---------|---------|
| C K | 8.46 | 18.07 |
| O K | 34.31 | 55.01 |
| Na K | 0.39 | 0.43 |
| Si K | 0.20 | 0.18 |
| S K | 0.18 | 0.15 |
| Cl K | 0.65 | 0.47 |
| K K | 0.31 | 0.20 |
| Mn K | 0.58 | 0.27 |
| Fe K | 54.91 | 25.22 |
| Totals | 100.00 | |



INCA

Comment: campione ultrasuoni 1 K(alpha) ferro = 6.4KeV
Tempo acquisizione 120 s
Ingrandimento 3000x

Steriwave Startec Ltd

Physical and Chemical Analysis Ultrasonic Sample

Project 1

11/24/2009 4:24:37 PM



Spectrum processing :
Peak possibly omitted : 14,180 keV

Processing option : All elements analyzed (Normalised)
Number of iterations = 6

Standard :

C CaCO3 1-Jun-1999 12:00 AM
O SiO2 1-Jun-1999 12:00 AM
Na Albite 1-Jun-1999 12:00 AM
Si SiO2 1-Jun-1999 12:00 AM
S FeS2 1-Jun-1999 12:00 AM
Cl KCl 1-Jun-1999 12:00 AM
K MAD-10 Feldspar 1-Jun-1999 12:00 AM
Ca Wollastonite 1-Jun-1999 12:00 AM
Mn Mn 1-Jun-1999 12:00 AM
Fe Fe 1-Jun-1999 12:00 AM
Br KBr 1-Jun-1999 12:00 AM
W W 1-Jun-1999 12:00 AM

| Element | Weight% | Atomic% |
|---------|---------|---------|
| C K | 10.78 | 23.85 |
| O K | 27.77 | 46.13 |
| Na K | 0.70 | 0.81 |
| Si K | 0.27 | 0.25 |
| S K | 0.17 | 0.14 |
| Cl K | 0.76 | 0.57 |
| K K | 0.30 | 0.21 |
| Ca K | 0.23 | 0.16 |
| Mn K | 0.65 | 0.31 |
| Fe K | 57.54 | 27.38 |
| Br L | 0.36 | 0.12 |
| W M | 0.47 | 0.07 |
| Totals | 100.00 | |



Comment: campione ultrasuoni 2 K(alpha) ferro = 6.4KeV
Tempo acquisizione 120 s
Ingrandimento 3000x

Physical and Chemical Analysis Ultrasonic Sample

Project 1

11/24/2009 4:29:31 PM



Spectrum processing :
No peaks omitted

Processing option : All elements analyzed (Normalised)
Number of iterations = 7

Standard :

C CaCO₃ 1-Jun-1999 12:00 AM
O SiO₂ 1-Jun-1999 12:00 AM
Na Albite 1-Jun-1999 12:00 AM
Mg MgO 1-Jun-1999 12:00 AM
Al Al₂O₃ 1-Jun-1999 12:00 AM
Si SiO₂ 1-Jun-1999 12:00 AM
S FeS₂ 1-Jun-1999 12:00 AM
Cl KCl 1-Jun-1999 12:00 AM
K MAD-10 Feldspar 1-Jun-1999 12:00 AM
Ca Wollastonite 1-Jun-1999 12:00 AM
Mn Mn 1-Jun-1999 12:00 AM
Fe Fe 1-Jun-1999 12:00 AM
W W 1-Jun-1999 12:00 AM

| Element | Weight% | Atomic% |
|---------|---------|---------|
| C K | 19.80 | 36.90 |
| O K | 29.27 | 40.96 |
| Na K | 1.20 | 1.16 |
| Mg K | 0.19 | 0.17 |
| Al K | 0.53 | 0.44 |
| Si K | 0.49 | 0.39 |
| S K | 0.27 | 0.19 |
| Cl K | 1.61 | 1.02 |
| K K | 0.54 | 0.31 |
| Ca K | 0.68 | 0.38 |
| Mn K | 0.47 | 0.19 |
| Fe K | 44.45 | 17.82 |
| W M | 0.50 | 0.06 |
| Totals | 100.00 | |



17


Comment: campione ultrasuoni 3 K(alpha) ferro = 6.4KeV
Tempo acquisizione 120 s
Ingrandimento 3000x

Inca

Steriwave Startec Ltd


Physical and Chemical Analysis

Normal analysed zone



Physical and Chemical Analysis

Ultrasonic analysed zone



Neutron Probe BTI



Bubble Technology Industries N-PROBE is a revolutionary portable neutron scintillation spectrometer, designed to be used by non-specialists for measurement of low-intensity neutron doses in the mixed field environments often found in nuclear utilities, fuel storage areas, fuel and waste processing operations and military applications. It is compatible with the current generation of BTI MICROSPEC analyzers and shares the philosophy of spectral dosimetry with other BTI spectroscopic probes, where the dosimetric quantities are computed from the spectrum using appropriate fluence-dose conversion functions.

The N-Probe uses two separate detectors to cover the neutron energy range from thermal to 20 MeV. A NE213-type liquid scintillator is used to cover the energy region from thermal to 800 keV. A ^3He proportional counter based on the $^3\text{He}(n,p)$ T reaction is used to cover the energy region from thermal to 800 keV. Sophisticated pulse-shape discrimination circuitry distinguishes unwanted gamma-ray signals from the desired neutron signals. The two detectors work in unison and pulse-height distributions from both are shown during data collection. With one command from the keyboard, these distributions are merged and processed automatically to yield the desired neutron spectrum.

● ● ● | HDS-100G/GN Mirion-MPG



HDS-100G/GN



Strumento portatile di ricerca
radiazioni gamma e neutroni


• Ricerca e rilevazione di materiali radioattivi gamma
e neutroni • Soddisfa tutte le richieste in materia di controlli di
sicurezza ambientale

R-1-S Reactor demo model



Outside view of the containment chamber and control system

R-1-S Reactor demo model



Inside view of the containment chamber and control system

● ● ● R-1-S Reactor demo model



Inside view of the containment chamber and control system



Contact

Steriwave Startec Ltd

STERIWAVE STARTEC LTD

English JV Office: Airport House Purley Way
Croydon Surrey CROOXZ
+44 7894 262399
info@steriwavestartec.com